
## NOTES:


1. SUBSTRATE: N-BK7

- 2. S2 TO BE PARALLEL WITH \$1 TO <1 ARCMIN
- 3. COATING (APPLY ACROSS CLEAR APERTURE) S1 & S2: RABS ≤0.25% @ 880nm RAVG ≤1.25% @ 400 - 870 nm RAVG ≤1.25% @ 890 - 1000nm



- 5. Rohs Compliant
- 6. SURFACE FLATNESS (P-V):  $\lambda/4$  (per inch over clear aperture)





## PARTS TO THIS DRAWING

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

> SHEET 1 OF 1

	\$1	\$2
SHAPE	PLANO	PLANO
SURFACE QUALITY	60-40	60-40
CLEAR APERTURE	Ø9.00	Ø9.00
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED

			Edmund Optic	;S®
HIRD ANGLE PROJECTION	$\phi$	TITLE	10mm DIA., 2mm THICK, VIS-NIR COA λ/4 N-BK7 WINDOW	TED
ALL DIMS IN	mm	DWG NO	45657	SHEET 1 OF